

Weightron Bilanciai Ltd,
Titan Works
Bridge Way off Broombank Road,
Chesterfield Trading Estate,
Chesterfield S41 9QJ, UK
Tel: 01246 260062
Fax: 01246 260844
www.weightron.com



Weightron Bilanciai Ltd,
Unit 4b, Block 4
Westpark, Chelston
Wellington TA21 9AD, UK
Tel: 01823 662355
Fax: 01823 662017
www.weightron.com



CALIBRATION CERTIFICATE

This Calibration Certificate is issued in accordance with the UKWF Code of Practice for the calibration of Non Automatic Weighing Machines. All weights used are traceable to NPL

Customer

EGGEN UK LTD
CROSS GREEN VALE
CROSS GREEN INDUSTRIAL
LEADS
LSA055

LOCATION
IN/OUT

Machine Details

Cust ref	88450
Make	WEIGHTRON
Model	EV22
Ser No	90177
Capacity	50000kg
e/d	20g
Working cap	

Test Weights

Set No.	Cert No.	Date Cal
1	TA10472	6-6-18

Cert Interval	ANNUAL
Next Cert	DEC 2019

Tolerance +/-	KL
0-10000	20
10000-40000	40
40000-50000	60

Comments ALL WEIGHTS IN KL
FOUR LITRE - 4320
TEST UNIT - 19600
RECALIBRATION AFTER MOVE

2	3
1	4

AS FOUND LINEARITY & HYSTERESIS

Load	Loading	Dev	Unloading	Dev
0	0	0	0	0
400	400	0	400	0
10000	10000	0	10000	0
20000	19980	-20	19980	-20
25600	25580	-20	25580	-20
29600	29580	-20	29580	-20
35600	35560	-40	35560	-40
39600	39560	-40	39560	-40
44400	44360	-40	44360	-40

DEFINITIVE LINEARITY & HYSTERESIS

Load	Loading	Dev	Unloading	Dev
0	0	0	0	0
400	400	0	400	0
10000	10000	0	10000	0
20000	20000	0	20000	0
25600	25600	0	25600	0
29600	29600	0	29600	0
35600	35600	0	35600	0
39600	39600	0	39600	0
44400	44400	0	44400	0

Repeatability

Load	Ind	Ind	Ind
44400	44360	44360	44360

Repeatability

Load	Ind	Ind	Ind
44400	44400	44400	44400

Eccentric Load - 16000

Reading	Dev	Reading	Dev
1	15980	2	15980
3	15980	4	15980
5		6	
7		8	

Eccentric Load - 16000

Reading	Dev	Reading	Dev
1	16000	2	16000
3	16000	4	16000
5		6	
7		8	

Signature Name Chalc McCusson Date 20-12-18